



深圳讯丰通医疗股份有限公司
Shenzhen XFT Medical Limited

Biocompatibility Evaluation

受控正本

| | |
|-----------------------|--|
| Product Name: | Nerve and Muscle Stimulator |
| Product Model | XFT-2003EA |
| Document No.: | ENNE-2003EA-147 |
| Version: | C |
| Document Type | Development Document |
| Confidentiality Level | Confidential |
| Project Code | RD48 |
| Printed Color | <input checked="" type="checkbox"/> Black <input type="checkbox"/> Colour |
| Issue Dept. | <input checked="" type="checkbox"/> R&D Dept. <input type="checkbox"/> Quality & Regulatory Dept. <input type="checkbox"/> Production Management Dept. <input type="checkbox"/> Marketing Center <input type="checkbox"/> QC Dept. <input type="checkbox"/> Purchasing Dept. <input type="checkbox"/> Engineering Dept. |

| | | |
|--------------|--|-----------------|
| Prepared by: | | Date: 2020.7.28 |
| Reviewed by: | | Date: 2020.7.28 |
| Approved by | | Date: 2020.7.28 |

Revision Record

[illegible]

1 The product name and model

Trade name: Hand Rehab System

Product name: Nerve and Muscle Stimulator

Model: XFT-2003EA


2 The purpose of evaluation

Nerve and Muscle Stimulator (Model: XFT-2003EA) is the final product which need perform as intended and be safe for human use, especially including a biological evaluation according to FDA guidance- Use of International Standard ISO 10993-1, "Biological evaluation of medical devices-Part 1: Evaluation and testing within a risk management process.

3 Evaluation of medical device

3.1 Materials/components in contact with the human body

3.1.1 Illustration of accessible components



3.1.2 List of patient contact components

| Model# | Components | Patient contact material | Body Contact Location | Contact duration |
|------------|--------------------------|---|---|------------------|
| XFT-2003EA | Top cover | ABS920 | Indirect contact, patient hand may contact it during operation | More than 30days |
| | Bottom cover | PC+SUS316 | Indirect contact, patient hand will contact it during operation | More than 30days |
| | Metal electrode | SUS316 | Direct contact, patient hand will contact it during operation | More than 30days |
| | Wristband | ABS920+TPE(534 U)+SUS316 | Direct contact, patient hand will contact it during operation | More than 30days |
| | Hook & Loop (magic tape) | Imported nylon non-scratch 6210 Velcro, Imported nylon non-scratch 6210 Velcro | Indirect contact, patient hand may contact it during fasten the cuff. | More than 30days |
| | Display panel | PC | Indirect contact, patient hand will contact it during operation | More than 30days |

Note: The above materials are very common in daily life, these materials have the characteristic of has a high performance due to its high resistance to mechanical damage and protection from environmental factors. So, they have been widely used an applied in the medical industry recent years.

- 3.2 According to the description of section 3.1.1 and 3.1.2 in this document, the device can be considered as a surface device and contact with the intact skin of the body, the contact duration is less than 24 hours. Therefore, the device is categorized as "A" in according to Table A.1 of FDA guidance- Use of International Standard ISO 10993-1, "Biological evaluation of medical devices-Part 1: Evaluation and testing within a risk management process. The risk of adverse reaction such as in vitro cytotoxicity, skin irritation and skin sensitization should be considered.

Table A.1: Biocompatibility Evaluation Endpoints

| Medical device categorization by | | | Biological effect | | | | | | | | | | | | |
|----------------------------------|---------------------------------|--|-------------------|---------------|---|-------------------------|--------------------------------|------------------------------|--------------|--------------|-------------------|------------------|-----------------|--------------------------------------|--------------|
| Nature of Body Contact | | Contact Duration A – limited (≤24 h) B – prolonged (>24 h to 30 d) C – permanent (> 30 d) | Cytotoxicity | Sensitization | Irritation or Intracutaneous Reactivity | Acute Systemic Toxicity | Material-Mediated Pyrogenicity | Subacute/Subchronic Toxicity | Genotoxicity | Implantation | Hemocompatibility | Chronic Toxicity | Carcinogenicity | Reproductive/Developmental Toxicity# | Degradation@ |
| Category | Contact | | | | | | | | | | | | | | |
| Surface device | Intact skin | A | X | X | X | | | | | | | | | | |
| | | B | X | X | X | | | | | | | | | | |
| | | C | X | X | X | | | | | | | | | | |
| | Mucosal membrane | A | X | X | X | | | | | | | | | | |
| | | B | X | X | X | O | O | O | | O | | | | | |
| | | C | X | X | X | O | O | X | X | O | | | O | | |
| | Breached or compromised surface | A | X | X | X | O | O | O | | O | | | | | |
| | | B | X | X | X | O | O | O | | O | | | | | |
| | | C | X | X | X | O | O | X | X | O | | | O | O | |
| External communicating device | Blood path, indirect | A | X | X | X | X | O | | | | X | | | | |
| | | B | X | X | X | X | O | O | | | X | | | | |
| | | C | X | X | O | X | O | X | X | O | X | O | O | | |

3.3 Evaluation endpoints for consideration

| Model | Components | Material | Corresponding Bio-compatibility Test | | |
|----------------|--------------------------|---|---|---|---|
| | | | In Vitro Cytotoxicity | Skin Sensitization | Skin Irritation |
| XFT-2003E A | Top cover | ABS920 | ISO10993-5: 2009 Biological evaluation of medical devices -Part 5: Tests for in vitro cytotoxicity | ISO10993-10: 2010 Biological evaluation of medical devices -Part 10: Tests for irritation and delayed-type hypersensitivity | ISO10993-10: 2010 Biological evaluation of medical devices -Part 10: Tests for irritation and delayed-type hypersensitivity |
| | Bottom cover | PC+SUS316 | | | |
| | Metal electrode | SUS316 | | | |
| | Wrist band | ABS920+TPE(534 U)+SUS316 | | | |
| | Hook & Loop (magic tape) | Imported nylon non-scratch 6210 Velcro, Imported nylon non-scratch 6210 Velcro | | | |
| | Display panel | PC | | | |

4 Biological Safety Testing Results

| Model | Test Item | Test Report Number | Test Result |
|------------|-----------------------|--|--|
| XFT-2003EA | In Vitro Cytotoxicity | SDWH-M201900697-1 (E) (007_ In Vitro Cytotoxicity Test) | The result showed that the test article has no potential toxicity |
| | Skin Sensitization | SDWH-M201900697-3 (E) (007_ Skin Sensitization Test Using Sesame Oil Extract); SDWH-M201900697-2 (E) (007_ Skin Sensitization Test Using 0.9% Sodium Chloride Injection Extract) | The results showed that the test article has no significant evidence of causing skin sensitization |
| | Skin Irritation | SDWH-M201900697-5 (E) Amd01(E) (007_ Skin Irritation Test Using Sesame Oil Extract); SDWH-M201900697-4 (E) (007_ Skin Irritation Test Using 0.9% Sodium Chloride Injection Extract) | The test results showed that the response of the test article extract was categorized as negligible under the test condition |

5 Conclusion

The biocompatibility test result shows that the materials which used for the device did not have any potential toxicity, skin sensitization and skin irritation. In



conclusion, the biocompatibility of the Nerve and Muscle Stimulator (Model: XFT-2003EA,) conforms to the intended use and regulation requirement.