

## **EC DECLARATION OF CONFORMITY**

Manufacturer: NAQI nv

Stadsbeemd 1037 B-3545 Halen Belgium

Devices: Ultrason gel

Basic UDI-DI: 542502989USUN

Manufacturer SRN: BE-MF-000001241

Device codes:

Device name	Device code	UDI-DI (GTIN)
A Connect Gel 500 ml	CS-US500	05425029892415

Intended purpose: used on the skin as contact gel for the transmission of ultrasound

waves in electrotherapy or ultrasound examination

Risk class: Class I Medical device based on rule 1 MDR 2017/745 Annex VIII

We, NAQI NV, hereby declare and confirm, under our sole responsibility as manufacturer, that the Ultrason gel devices covered by this declaration are in conformity with the requirements of EU regulation 2017/745.

Date : 01/10/2021

Name and function: E. M. Geyskens CEO of NAQI

Signature